# Visium CytAssist Spatial Gene and Protein Expression –

Deparaffinization, Decrosslinking, Immunofluorescence Staining & Imaging for FFPE Tissue

#### Introduction

The Visium CytAssist Spatial Gene and Protein Expression assay is designed to analyze mRNA and protein expression in tissue sections derived from formalin fixed & paraffin embedded (FFPE) tissue samples. A single CytAssist run accommodates up to two tissue slides (tissue placed on a standard glass slide) as sample input. This protocol outlines deparaffinization, decrosslinking, immunofluorescence (IF) staining, and imaging of FFPE tissue for use with 10x Genomics Visium CytAssist Spatial Gene Expression assay.

#### **Additional Guidance**

Consult the Visium CytAssist Spatial Gene and Protein Expression - FFPE Tissue Preparation Guide (Document CG000660) for complete information on sectioning FFPE tissue blocks and placing sections on slides. Consult the Visium CytAssist Spatial Applications Imaging Guidelines Technical Note (Document CG000521) to verify imaging settings before starting this Demonstrated Protocol. After completing this Demonstrated Protocol, proceed with the Visium CytAssist Spatial Gene and Protein Expression - User Guide (CG000494).

#### **Contents**

- 2 Reagent Kits
- 5 Recommended Thermal Cyclers
- **6** Workflow Overview
- 7 Specific Reagents & Consumables
- 9 Tips & Best Practices
- 25 1. Deparaffinization, Decrosslinking & IF Staining
- 26 1.1 Preparation Buffers
- 29 Protocol Overview
- 33 1.2 Deparaffinization
- 35 1.3 Decrosslinking
- 36 1.4 Immunofluorescence Staining
- 43 1.5 Coverslipping

- 44 2. Tissue Imaging
- 44 2.1 Imaging System Recommendations
- 45 2.2 Imaging
- 45 2.3 Coverslip Removal
- 46 Troubleshooting



## **Reagent Kits**

## Visium CytAssist Spatial Gene and Protein Expression Reagent Kits

Refer to SDS for handling and disposal information

Visium CytAssist Slide and Cassettes, 6.5 mm, 2 rxns PN-1000519

Visium CytAssist Slide and Cassettes, 6.5 mm 2 rxns PN-1000519			
Store at ambient temperature	#	PN	
Visium Cassette, 8 port	1	3000811	
Visium CytAssist moveable gasket small (preassembled with translator)	2	3000814	
Visium CytAssist moveable translator (preassembled with gasket)	2	3000816	
Visium CytAssist moveable Cassette, frame	2	3000813	
Visium CytAssist Slide Seals, 40 pack	1	2000284	
Visium CytAssist Spatial Gene Expression Slide v2, 6.5 mm	1	2000549	
			10

10

Visium CytAssist Slide and Cassettes, 11 mm, 2 rxns PN-1000518


Visium CytAssist Slide and Cassettes, 11 mm 2 rxns			
PN-1000518			
LIV-1000319			
Store at ambient temperature	#	PN	
Visium Cassette, 2 port	1	3000812	
Visium CytAssist moveable gasket large	2	3000815	
Visium CytAssist moveable Cassette, frame	2	3000813	
Visium CytAssist Slide Seals, 40 pack	1	2000284	
Visium CytAssist Spatial Gene Expression Slide v2, 11 mm	1	2000701	
			10

## **Reagent Kits**

#### Visium CytAssist Spatial Gene and Protein Expression Reagent Kits

Refer to SDS for handling and disposal information

## Visium FFPE Reagent Kit v2 - Small PN-1000436



<sup>\*</sup>Only these reagents are used in this Demonstrated Protocol.

Enough reagent is provided for processing two 6.5 mm slides or one 11 mm slide. Information on how much reagent is used for antibody optimization is provided on page 22. Enough reagent is provided for at least 13 antibody optimizations if using 6.5 mm slides and 6 optimizations if using 11 mm slides.

<sup>\*\*</sup>These tubes may not be included in the kit. They are not used in this assay.

## **Reagent Kits**

## Visium CytAssist Spatial Gene and Protein Expression Reagent Kits

- Refer to SDS for handling and disposal information.
- Visium and Tissue Slide Cassettes are single use.



- Additional Visium Slide Cassettes (PN-1000469) should be purchased for antibody optimization.
- Additional Tissue Slide Cassettes (PN-1000471, PN-1000472) must be purchased for each processed section.

Visium 6.5 mm Slide Cassette, 4 pk PN-1000469

Visium 6.5 mm Slide Cassette, 4 pk			
PN-1000469 Store at ambient temperature	#	PN	
Visium Cassette, 8 port	4	3000811	
			10x

Visium CytAssist Tissue Slide Cassette, 6.5 mm, 4 pk PN-1000471

Visium CytAssist Tissue Slide Cassette, 6.5 mm 4 pk PN-1000471	"		
Store at ambient temperature	#	PN	
Visium CytAssist moveable gasket small (preassembled with translator)	4	3000814	
Visium CytAssist moveable translator (preassembled with gasket)	4	3000816	
Visium CytAssist moveable Cassette, frame	4	3000813	
			10×

Visium CytAssist Tissue Slide Cassette, 11 mm, 4 pk PN-1000472

Visium CytAssist Tissue Slide Cassette, 11 m 4 pk PN-1000472	m		
Store at ambient temperature	#	PN	
Visium CytAssist moveable gasket large	4	3000815	
Visium CytAssist moveable Cassette, frame	4	3000813	
			10x

#### **10x Genomics Accessories**

Product	#	Kit and Part Number	Part Number (Item)
Low Profile Plate Insert	2	Visium CytAssist Reagent	3000823
10x Magnetic Separator	1	Accessory Kit: 1000499	230003

## **Recommended Thermal Cyclers**

Supplier	Description	Part Number
Bio-Rad	C1000 Touch Thermal Cycler with 96-Deep Well Reaction Module	1851197
Eppendorf	MasterCycler Pro (discontinued)	North America 950030010 International 6321 000.019
	MasterCycler X50s	North America 6311000010
Thermo Fisher Scientific	Veriti 96-Well Thermal Cycler (discontinued)	4375786
	VeritiPro Thermal Cycler, 96-well*	A48141
Analytik Jena	Biometra TAdvanced 96 SG with 96-well block (silver, 0.2 mL) and gradient function	846-x-070-241

\*Requires firmware 1.2.0 or later. Open lid slowly to avoid disturbing the cassette. After incubation, if the cassette becomes stuck to the lid, remove carefully. When working with tubes, VeritiPro requires blue MicroAmp 96-Well Tray/Retainer Set for Veriti Systems (PN-4381850), with top piece removed.




The following thermal cyclers should have their ramp rates adjusted for all steps:

- Eppendorf MasterCycler X50s: 3°C/sec heating and 2°C/sec cooling
- Analytik Jena Biometra TAdvanced: 2°C/sec heating and cooling


Thermal cycler must be able to accommodate the Low Profile Plate Insert (also referred to as the Low Profile Thermocycler Adapter):

- Well depth: 4.5 mm
- Distance between block and heated lid: 12 mm
- Reaction block dimensions 95.5 x 73 mm

The Low Profile Thermocycler Adapter may be difficult to remove from some thermal cyclers. If this is the case, turn off the thermal cycler and place a -20°C cooling pad on top of the Low Profile Thermocycler Adapter for a few minutes prior to attempting removal. Do not handle the Low Profile Thermocycler Adapter while it is hot.



#### **Workflow Overview**



Visit the 10x Genomics Support website for the most updated documentation.

## **Specific Reagents & Consumables**

For each item, a number of vendor options are listed. Choose item based on availability and preference. **Substituting materials may adversely affect system performance**.

Item	Alternatives/Options	Vendor	Part Number
Xylene	Xylene, Reagent Grade	Millipore Sigma	214736
	Xylene, Histological Grade	Millipore Sigma	534056
Glycerol	Glycerol	Acros Organics	327255000
	Glycerol Solution	Millipore Sigma	49781
Ethanol	Ethyl Alcohol, 200 Proof	Millipore Sigma	E7023
	Ethanol absolute ≥99.5%, TechniSolv, pure (Europe Only)	VWR	83813.360DP
PBS	PBS - Phosphate Buffered Saline (10X) pH 7.4, RNase-free	Thermo Fisher Scientific	AM9624
SSC	SSC Buffer 20x Concentrate	Millipore Sigma	S66391L
10% Tween-20	Tween 20 Surfact-Amps Detergent Solution (10% solution)	Thermo Fisher Scientific	28320
Nuclease-free water	Nuclease-free Water (not DEPC-Treated)	Thermo Fisher Scientific	AM9937
DAPI	DAPI Solution (1 mg/ml)	Thermo Fisher Scientific	62248
Antibodies	-	-	-
RNase inhibitor	Protector RNase Inhibitor	Millipore Sigma	3335399001
Mounting medium	SlowFade Diamond Antifade Mountant	Thermo Fisher Scientific	S36967
	SlowFade Gold Antifade Mountant	Thermo Fisher Scientific	S36966
Section dryer oven	Epredia High Capacity Section Dryer	Fisher Scientific	A84600051
Slide holders	Slide Holders, 24-place	VWR	25608-868
Coverslips	Fisherbrand Cover Glasses: Rectangles	Fisher Scientific	12-544-EP
	Cover Glasses, Rectangular	VWR	16004-322
Staining jar/dishes	Coplin Jar	VWR	100500-232
	Staining Dishes	VWR	25608-906
Green Marker, Optional, if annotating slide	Sharpie Argyle Green Permanent Marker Or any equivalent hue	Sharpie	1785396
Autofluorescence Quencher	TrueBlack Lipofuscin Autofluorescence Quencher Optional	Biotium	23007

## **Specific Reagents & Consumables**

For each item, a number of vendor options are listed. Choose item based on availability and preference. **Substituting materials may adversely affect system performance**.

Pipet-Lite Multi Pipette L8-200XLS+	Rainin	17013805
Pipet-Lite LTS Pipette L-2XLS+	Rainin	17014393
Pipet-Lite LTS Pipette L-10XLS+	Rainin	17014388
Pipet-Lite LTS Pipette L-20XLS+	Rainin	17014392
Pipet-Lite LTS Pipette L-100XLS+	Rainin	17014384
Pipet-Lite LTS Pipette L-200XLS+	Rainin	17014391
Pipet-Lite LTS Pipette L-1000XLS+	Rainin	17014382
Tips RT LTS 200UL FLW	Rainin	30389241
Tips LTS 200UL Filter RT-L200 FLR	Rainin	30389240
Tips LTS 1ML Filter RT-L1000 FLR	Rainin	30389213
Tips LTS 20UL Filter RT-L20 FLR	Rainin	30389226
	-	-
er,	-	-
	Pipet-Lite LTS Pipette L-2XLS+  Pipet-Lite LTS Pipette L-10XLS+  Pipet-Lite LTS Pipette L-20XLS+  Pipet-Lite LTS Pipette L-100XLS+  Pipet-Lite LTS Pipette L-200XLS+  Pipet-Lite LTS Pipette L-1000XLS+  Tips RT LTS 200UL FLW  Tips LTS 200UL Filter RT-L200 FLR  Tips LTS 1ML Filter RT-L1000 FLR  Tips LTS 20UL Filter RT-L20 FLR	Pipet-Lite LTS Pipette L-2XLS+ Rainin  Pipet-Lite LTS Pipette L-10XLS+ Rainin  Pipet-Lite LTS Pipette L-20XLS+ Rainin  Pipet-Lite LTS Pipette L-100XLS+ Rainin  Pipet-Lite LTS Pipette L-200XLS+ Rainin  Pipet-Lite LTS Pipette L-1000XLS+ Rainin  Pipet-Lite LTS Pipette L-1000XLS+ Rainin  Tips RT LTS 200UL FLW Rainin  Tips LTS 200UL Filter RT-L200 FLR Rainin  Tips LTS 1ML Filter RT-L1000 FLR Rainin  Tips LTS 20UL Filter RT-L20 FLR Rainin  Rainin

# **Tips & Best Practices**



Tips & Best Practices section includes additional guidance



Signifies critical step requiring accurate execution



Troubleshooting section includes additional guidance

## **General Reagent Handling**


• Thoroughly mix reagents before use.

## **Pipette Calibration**

• Follow manufacturer's calibration and maintenance schedules.

## **Slide Handling**

- Always wear gloves when handling slides.
- DO NOT touch the tissue sections on the slide.
- Minimize exposure of the slides to sources of particles and fibers.
- When immersing slides in deparaffinization solutions and water, ensure that the tissue sections are completely submerged. DO NOT submerge slide label.
- Keep the slide flat on the bench when adding reagents to the active surface.
- Ensure that no absorbent surface is in contact with the reagents on the slide during incubation.
- DO NOT use a hydrophobic barrier (such as one applied by a PAP pen) on slides. These barriers may affect assay performance.
- Consider tracing the Tissue Slide Cassette gasket onto the back of the tissue slide to assist in alignment on the CytAssist instrument.
  - Markings may cause automatic tissue registration to fail, resulting in a need for manual registration.
  - If leaving a mark is desired, a green colored marker (Sharpie, PN-1785396, or any equivalent hue) has the least impact on the automatic tissue registration process.



#### Slide Incubation Guidance

## Incubation at a specified temperature

#### **Incubation using a Section Dryer Oven:**

- Place the slides in a slide drying rack sideways to minimize paraffin wax entering neighboring tissue.
- Close door when incubating the slide in the oven.



#### **Incubation using a Thermal Cycler:**

- Position a Low Profile Plate Insert (also referred to as Low Profile Thermocycler Adapter) on a thermal cycler that is set at the incubation temperature.
- · Move Low Profile Thermocycler Adapter back and forth to ensure that it is seated properly.
- Ensure that the Low Profile Thermocycler Adapter is in contact with the thermal cycler surface uniformly.
- When incubating a slide encased in a cassette, place the assembled unit on the Low Profile Thermocycler Adapter with the wells facing up.
- The cassette should sit flat on the Low Profile Thermocycler Adapter.
- Cassettes should always be sealed when in the Low Profile Thermocycler Adapter.
- Tighten the thermal cycler lid, but do not overtighten.
- · After removing the cassette from the thermal cycler, ensure that volume removed from the well is as expected. Volumes lower than what are expected may indicate an improper seal, resulting in evaporation.
- Allow Low Profile Thermocycler Adapter to cool before removing it from the thermal cycler.



#### **Slide Incubation Guidance**

## Incubation at room temperature


• Place the slide in the cassette on a flat, clean, nonabsorbent work surface.


## **Visium CytAssist Tissue Slide Cassette**

- · The Visium CytAssist Tissue Slide Cassette encases the slide and creates a leakproof well for adding reagents on tissue slides.
- The cassette is a single use item.
- · Gaskets are adjustable to ensure that the tissue section or area of interest is encased in a well.
- Refer to Visium CytAssist Tissue Slide Cassette Assembly & Removal instructions for details.



- · Before use, inspect the moveable gasket to ensure that the gasket perimeter and corners are free of excess silicone before assembly.
- Excess silicone should be safely removed with forceps or a pipette tip before assembly.
- · Assembly should occur against a white background for easy tissue visualization during alignment.
- Practice assembly with a blank slide (75 x 25 x 1 mm).
- · Place slides in the cassette only when specified.





## **Visium CytAssist Tissue Slide Cassette Assembly**



Wear fresh gloves while assembling Tissue Slide Cassette.



Exercise caution when handling slide edges to prevent injury.


Break cassette into two halves by bending each half at the hinge until they snap apart



The 6.5 mm gasket can be adjusted horizontally and rotated in two directions (180°) while 11 mm gasket can be rotated in four directions (90°). Determine the appropriate configuration that allows the gasket to encompass the tissue area of interest.



Gently place top half of cassette over bottom half. DO NOT assemble together until Step 7.



Apply even pressure on top of cassette until it clicks shut. Verify that clip is completely secured over hinges.



Place tissue slide into lower half of cassette with tissue facing up



Securely combine gasket with top half of cassette until the gasket snaps into place.



Adjust gasket such that gasket is over the tissue area of interest. The 6.5 mm gasket can be adjusted horizontally as well as vertically.




Turn cassette over and verify tissue area of interest is within gasket. DO NOT move gasket once cassette is closed. If necessary, open cassette and recenter gasket.



## **Visium CytAssist Tissue Slide Cassette Removal**

Pull clip up to detach upper and lower halves of cassette



Open cassette by continuing to lift clip upward. If slide sticks to gasket, continue to apply even upward pressure to separate slide from gasket




Hold slide by the label and lift slide out from lower half




#### **Visium Cassette**

- The Visium Cassette encases the slide and creates leakproof wells for reagent addition.
- The Visium Cassette is a single use item.
- Place the slides in the Visium Cassette only when specified.
- · See Visium Cassette Assembly & Removal instructions for details on assembly and removal.
- Ensure that the Visium Cassette and slide facing gasket are free of debris before assembly.
- Excess silicone in exhaust channels or within wells should be safely removed with forceps or a pipette tip before cassette assembly.
- · Failure to remove excess silicone may prevent adequate venting of the cassette during heated incubation steps.
- If the gasket appears warped, the Visium Cassette is still safe to use as long as the cassette can fully close.
- Visually inspect the gasket to ensure it is seated properly.
- Practice assembly with a blank slide (75 x 25 x 1 mm).
- Applying excessive force to the slide and/or cassette may cause the slide to break.



## **Visium Cassette**






## **Visium Cassette Assembly**

- Ensure that the surface of the cassette is dry.
- Cassette may also be assembled in the hand for comfort.




Break cassette into two halves by bending each half at the hinge until they snap apart



Place upper and lower halves of cassette, and Visium slide, directed upward on bench



Place Visium slide with active surface facing upwards into lower half of cassette; ensure label is toward hinges




Press slide down into grooves of the bottom half of the cassette until it sits firmly in place



Secure outer clips of top half with outer tab of bottom half



Press firmly on top of cassette until it clicks



Slides in images are representative.



## **Visium Cassette Removal**




Assembly and removal steps apply to both 6.5 and 11 mm Cassettes.

Pull inner clip up from inner tab to detach upper and lower halves of cassette



Open cassette by continuing to lift inner clip upward



Lift slide out from lower half




Slides in images are representative.

## **Reagent Addition to Wells**

- Place the assembled slide in the cassette flat on a clean work surface.
- · Dispense reagents along the side of the wells without touching the tissue sections and without introducing bubbles.



- Always cover the tissue section completely when adding reagents to the well. A gentle tap may help spread the reagent more evenly.
- Ensure that no bubbles are introduced in the process.



## **Reagent Removal from Wells**

- Slightly tilt the cassette while removing the reagent.
- Place the pipette tip to the bottom edge of the wells.
- Remove reagents along the side of the wells. Do not touch the tissue sections or introduce bubbles.
- Ensure that no bubbles are introduced in the process.
- · Remove all liquid from the wells in each step. To ensure complete removal, check the bottom of the well by tilting the cassette slightly. A meniscus at the bottom of the well will indicate the presence of liquid in the well. Repeat removal steps until no reagent remains.



## Visium CytAssist Slide Seal Application & Removal

## **Application**

- If applying a Visium Slide Seal to a Tissue Slide Cassette, the seal must be cut in half lengthwise.
- Six pre-cut seals per tissue section are needed for the entire Visium CytAssist Spatial Gene and Protein Expression assay.
- Cut the seal as shown in the image below. Ensure scissors are cleaned with a RNase decontamination solution.
- Place the CytAssist Tissue Slide Cassette flat on a clean work surface.
- Remove the back of the adhesive Visium Slide Seal.
- · Align the Visium Slide Seal with the surface of the cassette and apply while firmly holding the cassette with one hand.
- Press on the Visium Slide Seal to ensure uniform adhesion.

#### Removal

- Place the CytAssist Tissue Slide Cassette flat on a clean work surface.
- Pull on the Visium Slide Seal from the edge while firmly holding the cassette.
- Ensure that no liquid splashes out of the wells.



## Visium CytAssist Tested Slides

The following slides have been tested for use with the Visium CytAssist instrument.


Item	Length (mm)	Width (mm)	Thickness (mm)	Ground Corners
Epredia Shandon ColorFrost Plus	75.0	25.0	1.0	No
Fisherbrand SuperFrost Plus	75.0	25.0	1.0	Available as either
Sigma-Aldrich Poly Prep Slides	75.0	25.0	1.0	No
VWR SuperFrost Plus Micro Slide, Premium	75.0	25.0	1.0	No

If unsure of slide part number, refer to the "generic" slide diagram below for general guidance (images not to scale). Diagrams for verifying that tissue sections are placed in the allowable area can also be found in the Visium CytAssist Quick Reference Cards - Accessory Kit (Document CG000548). The diagrams demonstrate allowable areas that are far enough away from frosted sections to not interfere with gasket closure during the CytAssist assay. Frosted sections include the opaque area of the slide as well as any etching or writing on the slide.



While slides are specified as being 25 mm x 75 mm, manufacturing tolerances may lead to dimensions that are too small or large to be compatible with 10x Genomics products. Tissue slide dimensions must be within 24.8 mm - 25.3 mm in width and 74.4 mm - 76.2 mm in length to fit the Visium CytAssist Tissue Slide Cassette.

Minimum slide dimensions: 24.8 x 74.4 mm Maximum slide dimensions: 25.3 x 76.2 mm



## **Antibody Optimization**

Prior testing of antibodies is recommended on the same tissue block before performing immunofluorescence staining in combination with the Visium CytAssist Spatial Gene and Protein Expression workflow. Determination of the optimal antibody concentration is crucial for executing this protocol. Antibody optimization should be performed on the same area of interest in serial tissue sections.



This protocol uses the Visium Slide Cassettes, 6.5 mm; thus, tissue sections must fit within a 6.5 x 6.5 mm area. Additional Visium Slide Cassettes, 6.5 mm (PN-1000469) must be purchased for antibody optimization.

If larger tissue sections will be used in the main assay with the Visium CytAssist Spatial Gene and Protein Expression, 11 mm slides, identify a smaller, representative area of interest that can fit within the 6.5 x 6.5 mm areas in the cassette. Trim sections to avoid tissue overlap prior to placing on the blank slide. After placement, proceed with antibody optimization as described.

The tables below describe the **minimum** number of reactions (i.e. the number of wells in a Visium Cassette) available for antibody optimization, for a given reagent, while still leaving enough reagent for the IF staining protocol and main CytAssist assay. Ensure that enough reactions are available for the main assay after running optimization experiments.

The 6.5 mm slide table lists the number of optimization reactions that can be performed while still proceeding with four 6.5 mm samples in the IF protocol and main CytAssist assay. The 11 mm slide table lists the number of optimization reactions that can be performed while still proceeding with two 11 mm samples in the IF protocol and main CytAssist assay.


6.5 mm Slides			
Reagent	PN	Staining Method	# Reactions
Block and Stain Buffer	2000554	Fluorophore Conjugated Primary Antibodies	46
		Primary and Secondary Antibodies	31
Decrosslinking Buffer	2000566	-	13
11 mm Slides			
Reagent	PN	Staining Method	# Reactions
Block and Stain Buffer	2000554	Fluorophore Conjugated Primary Antibodies	23
		Primary and Secondary Antibodies	15
Decrosslinking Buffer	2000566	-	6

Optimal antibody concentrations for this Demonstrated Protocol may differ from other applications. Composition of reagents and buffers may also differ from other immunofluorescence applications.

When optimizing the antibody for a single tissue type, ensure that stained slides can be imaged according to the imaging guidelines listed in Visium CytAssist Spatial Applications Imaging Guidelines Technical Note (CG000521).

Below is a suggested optimization workflow. Perform antibody optimization according to preference, as long as tissues are tested using the staining protocol described in this document.


• Draw representative frames on the back of a 75 x 25 x 1 mm blank slide using the example slide layout.



- Tissue sections must be trimmed to fit within each 6.5 mm area on the template to prevent tissue overlap.
- · Place tissue sections in the frames on the front of the slide for compatibility with the Visium Cassette.
- Execute the Deparaffinization, Decrosslinking & Immunofluorescence Staining protocol using a range of antibody concentrations. A starting concentration of 0.01 ug/ul (0.7 µg/sample) is recommended.
- To reduce autofluorescence, TrueBlack Plus reagent may be added.
- Select the antibody concentration that results in the specific staining of desired cells, while minimizing nonspecific background staining.

Autofluorescence quenchers are added following immunofluorescence staining and may result in the reduction of fluorescence signal. Additional optimization and increase in antibody concentration may be required to visualize immunostaining.

An example dilution layout is provided below. DAPI and merged images are provided to show the presence of breast cancer tissue for each antibody dilution. Dilutions are of recombinant Anti-Vimentin antibody conjugated to Alexa Fluor 594 (BioLegend, PN677804, 0.5 mg/ml). A 1:100 dilution (0.35 µg/sample) was considered optimal in this example.



#### Example calculations:

- Stock antibody concentration is 0.5 μg/μl and desired concentration is 0.01 μg/μl (50X dilution).
- Stock antibody concentration is 0.5 μg/μl and desired concentration is 0.0025 μg/μl (200X dilution).
  - This calculation requires pipetting a small volume of antibody. Dilute the stock concentration first (for example, dilute 10X to 0.05 µg/µl) then proceed with the following reagent table:

Reagents	Volume (µI)
Block and Stain Buffer	17.5
RNase Inhibitor	3.5
Antibody	1.4
Nuclease-free Water	47.6
Total	70.0


# 1. Deparaffinization, Decrosslinking & Immunofluorescence (IF) Staining

#### 1.0 Overview

This chapter provides guidance on deparaffinization, decrosslinking, and immunofluorescence (IF) staining of tissue slides containing FFPE tissue sections that are dried overnight in a desiccator.



- Additional Tissue Slide Cassettes (PN-1000471, PN-1000472) must be purchased for each processed section.
- Ensure that microscope settings have been verified and imaging programs have been created before starting this program. Consult the Visium CytAssist Spatial Applications Imaging Guidelines Technical Note (CG000521) for more information.
- If staining using fluorophore conjugated primary antibodies, proceed directly to step 1.4A (Immunofluorescence staining - Fluorophore Conjugated Primary Antibodies) after completing step 1.3 (Decrosslinking).
- If staining using primary and secondary antibodies, proceed directly to step 1.4B (Immunofluorescence staining-Primary and Secondary Antibodies) after completing step 1.3 (Decrosslinking).



## 1.1 Preparation - Buffers

For	For Deparaffinization						
Pre	Prepare fresh weekly. Process two slides per jar.						
Items		Preparation & Handling					
☐ Xylene Label two coplin jars as Xylene Jar 1 and 2. Dispense 30 ml xylene in each.							
	100% Ethanol	Label two coplin jars as 100% Ethanol Jar 1 and 2. Dispense 30 ml 100% ethanol in each. Alternatively, use a 50-ml centrifuge tube or a beaker.					
	96% Ethanol	Label two coplin jars as 96% Ethanol Jar 1 and 2. Dispense 30 ml 96% ethanol in each. Alternatively, use a 50-ml centrifuge tube or a beaker.					
	70% Ethanol	Label one coplin jar as 70% Ethanol Jar. Dispense 30 ml 70% ethanol.  Alternatively, use a 50-ml centrifuge tube or a beaker.					
	Milli-Q or UltraPure Water	Label one coplin jar as Water Jar. Dispense 30 ml water. Alternatively, use a 50-ml centrifuge tube or a beaker.					



Alternatively, a slide staining dish can also be used in place of a coplin jar. Adjust the volumes of deparaffinization solutions and water, accordingly. Use xylene-resistant dishes, for immersion in xylene. Use xylene-resistant gloves or forceps for deparaffinization. Prepare fresh reagents every week.

For	For Decrosslinking						
Ite	ms	Preparation & Handling					
	1X PBS	Prepare 50 ml of 1X PBS using nuclease-free water.					
	2X SSC	Prepare 1 ml of 2X SSC using nuclease-free water.					
	70% Ethanol	Optional, if performing TrueBlack Quenching. Prepare 1 ml of 70% Ethanol by diluting stock ethanol with nuclease-free water.					
	Wash Buffer	Prepare Wash Buffer according to the appropriate table below, adding reagents in the order listed. Pipette mix 10x and centrifuge briefly. Maintain on ice.					

#### If using a fluorophore conjugated primary antibody:



6.5 mm Gaskets					
Wash Buffer	Stock	Final	1X (µl)	2X+ 10% (μl)	4X+ 10% (μl)
PBS	10X	1X	60.0	132.0	264.0
Tween-20	10%	0.4%	24.0	52.8	105.6
Nuclease-free Water	-		516.0	1,135.2	2,270.4
Total	-	-	600.0	1,320.0	2,640.0



11 mm Gaskets							
Wash Buffer	Stock	Final	1Χ (μl)	2X+ 10% (μl)	4X + 10% (μl)		
PBS	10X	1X	120.0	264.0	528.0		
Tween-20	10%	0.4%	48.0	105.6	211.2		
Nuclease-free Water	-		1,032.0	2,270.4	4,540.8		
Total	-	-	1,200.0	2,640.0	5,280.0		

## $\square$ Wash Buffer

#### If using a primary and secondary antibody:



6.5 mm Gaskets							
Wash Buffer	Stock	Final	1Χ (μl)	2X+ 10% (μl)	4X+ 10% (μl)		
PBS	10X	1X	120.0	264.0	528.0		
Tween-20	10%	0.4%	48.0	105.6	211.2		
Nuclease-free Water	-		1,032.0	2,270.4	4,540.8		
Total	-	-	1,200.0	2,640.0	5,280.0		



11 mm Gaskets							
Wash Buffer	Stock	Final	1Χ (μl)	2X+ 10% (μl)	4X + 10% (μl)		
PBS	10X	1X	240.0	528.0	1,056.0		
Tween-20	10%	0.4%	96.0	211.2	422.4		
Nuclease-free Water	-		2,064.0	4,540.8	9,081.6		
Total	-	-	2,400.0	5,280.0	10,560.0		

1X Blocking Buffer

Prepare Blocking Buffer on ice according to the appropriate table below, adding reagents in the order listed. Pipette mix 10x and centrifuge briefly. Maintain on ice.



6.5 mm Gaskets					
1X Blocking Buffer	Stock	Final	1X (µl)	2X+ 10% (μl)	4X+ 10% (μl)
Block and Stain Buffer	4X	1X	25.0	55.0	110.0
RNase Inhibitor	40X	1X	2.5	5.5	11.0
Nuclease-free Water	-	-	72.5	159.5	319.0
Total	-	-	100.0	220.0	440.0



11 mm Gaskets					
1X Blocking Buffer	Stock	Final	1X (µl)	2X+ 10% (μl)	4X + 10% (μl)
Block and Stain Buffer	4X	1X	50.0	110.0	220.0
RNase Inhibitor	40X	1X	5.0	11.0	22.0
Nuclease-free Water	-	-	145.0	319.0	638.0
Total	-	-	200.0	440.0	880.0

Diluted Decrosslinking Buffer

Thaw Decrosslinking Buffer at room temperature. Vortex and centrifuge briefly after preparing Diluted Decrosslinking Buffer.



6.5 mm Gaskets					
Diluted Decrosslinking Buffer	Stock	Final	1X (µl)	2X+ 10% (μl)	4X+ 10% (μl)
Decrosslinking Buffer	10X	1X	25	55	110
Nuclease-free Water	-		225	495	990
Total	-	-	250	550	1,100





11 mm Gaskets					
Diluted Decrosslinking Buffer	Stock	Final	1X (µl)	2X+ 10% (μl)	4X+ 10% (μl)
Decrosslinking Buffer	10X	1X	50	110	220
Nuclease-free Water	-		450	990	1,980
Total	-	-	500	1,100	2,200


For C	For Coverslipping						
Items		Preparation & Handling	g				
	Mounting Medium	Invert to mix. Briefly cer	ntrifuge to	o remove l	oubbles.		
		Mounting Medium	Stock	Final	1Χ (μl)	2X+ 15% (μl)	


Mounting Medium	Stock	Final	1Χ (μl)	2X+ 15% (μl)
Glycerol	100%	80%	80	184
SlowFade Diamond	100%	20%	20	46
Total	-	-	100	230

#### **Protocol Overview**









## 1.2 Deparaffinization

Deparaffinization steps should be performed in a fume hood due to the hazardous nature of xylene.

- **a.** Retrieve slides with tissue sections from the desiccator after overnight drying.
- b. Place slides in a Section Dryer Oven and incubate uncovered at 60°C for 2 h. Keep oven door closed during incubation.



Alternatively, place a Low Profile Thermocycler Adapter on a thermal cycler set at 60°C. Place slide on the Low Profile Thermocycler Adapter with the tissue side facing up and incubate 2 h at 60°C.



DO NOT close the thermal cycler lid.




**c.** Remove from the oven or thermal cycler and allow slides to cool down to room temperature.



TIPS When immersing slides in xylene, ensure that the tissue sections are completely submerged.

- **d.** Gently immerse slides 2x in Xylene Jar 1, then immerse and incubate for 10 min. Secure jar cap to prevent xylene loss.
- **e.** Gently immerse slides 2x in Xylene Jar 2, then immerse and incubate for **10 min**. Secure jar cap to prevent xylene loss.
- **f.** Gently immerse slides 2x in 100% Ethanol Jar 1, then immerse and incubate for 3 min.
- g. Gently immerse slides 2x in100% Ethanol Jar, then immerse and incubate for 3 min.



- **h.** Gently immerse slides 2x in 96% Ethanol Jar 1, then immerse and incubate for 3 min.
- i. Gently immerse slides 2x in 96% Ethanol Jar 2, then immerse and incubate for 3 min.
- **j.** Gently immerse slides 2x in 70% Ethanol Jar, then immerse and incubate for 3 min.
- k. Gently immerse slides 2x in Water Jar, then immerse and incubate for 20 sec.
- 1. Let the slides air dry and proceed to Decrosslinking.

## 1.3 Decrosslinking

- $\blacksquare$  denotes volumes for 6.5 mm gaskets and  $\blacktriangle$  denotes volumes for 11 mm gaskets
- a. Prepare the thermal cycler with the following incubation protocol and start the program.

Lid Temperature	Reaction Volume	Run Time
95°C	-	1 h
Step	Temperature	Time
Pre-equilibrate	95°C	Hold
Decrosslinking	95°C	00:60:00
Cooling	22°C	00:10:00
Hold	22°C	Hold

**b.** Place the slide in the Visium CytAssist Tissue Slide Cassette.



See Tips & Best Practices for assembly instructions. Practice assembly with a blank slide.

- **c.** Add ■150 µl or ▲300 µl Diluted Decrosslinking Buffer along the side of the wells.
- **d.** Remove all Decrosslinking Buffer from the wells.
- e. Add ■100 µl or ▲200 µl Diluted Decrosslinking Buffer along the side of the wells.
- **f.** Apply a pre-cut Visium Slide Seal on the cassette and place the cassette on the Low Profile Thermocycler Adapter at 95°C.
- g. Close the thermal cycler lid. Skip pre-equilibrate step and initiate Decrosslinking.
- **h.** After decrosslinking is complete, remove the cassette from the Low Profile Thermocycler Adapter and place on a flat, clean work surface.
- i. Peel back slide seal and using a pipette, remove all Decrosslinking Buffer from the well corners.
- j. Add ■150 µl or ▲300 µl 2X SSC along the side of the wells to uniformly cover the tissue sections, without introducing bubbles.
- **k.** Remove all 2X SSC from the wells.
- 1. Proceed **immediately** to appropriate Immunofluorescence Staining protocol.

## 1.4 Immunofluorescence Staining

Choose appropriate staining protocol depending upon the antibodies used.

■ denotes volumes for 6.5 mm gaskets and ▲ denotes volumes for 11 mm gaskets

## 1.4A Immunofluorescence Staining - Fluorophore Conjugated Primary Antibodies

a. Prepare Antibody Solution on ice, adding reagents in the order listed. Pipette mix 10x and centrifuge briefly. Maintain on ice.




proceed directly to

1.4B.



6.5 mm Gaskets					
Antibody Solution	Stock	Final	1Χ (μΙ)	2X+ 10% (μl)	4X+ 10% (μl)
Block and Stain Buffer	4X	1X	17.5	38.5	77.0
RNase Inhibitor	40U/µl	2U/µl	3.5	7.7	15.4
Antibody #1	Variable	Variable	Variable	Variable	Variable
Antibody #2 (Optional)	Variable	Variable	Variable	Variable	Variable
Antibody #3 (Optional)	Variable	Variable	Variable	Variable	Variable
DAPI Solution	100X	1X	0.7	1.5	3.1
Nuclease-free Water	Variable	Variable	Variable	Variable	Variable
Total	-	-	70.0	154.0	308.0



11 mm Gaskets					
Antibody Solution	Stock	Final	1Χ (μl)	2X+ 10% (µl)	4X + 10% (μl)
Block and Stain Buffer	4X	1X	35.0	77.0	154.0
RNase Inhibitor	40U/μl	2U/µl	7.0	15.4	30.8
Antibody #1	Variable	Variable	Variable	Variable	Variable
Antibody #2 (Optional)	Variable	Variable	Variable	Variable	Variable
Antibody #3 (Optional)	Variable	Variable	Variable	Variable	Variable
DAPI Solution	100X	1X	1.4	3.1	6.2
Nuclease-free Water	Variable	Variable	Variable	Variable	Variable
Total	-	-	140.0	308.0	616.0



Antibody dilution can change depending on the antibody, ranging from 1:50 down to 1:1000. Antibody volumes will depend on concentrations determined during antibody optimization. Add an appropriate volume of nuclease-free water based on the amount of added antibody to achieve the stated total volume.

**b.** Optional - Prepare 1X TrueBlack Solution according to the appropriate table, adding reagents in the order listed. Pipette mix 10x and centrifuge briefly. Maintain at room temperature.



If using primary and secondary antibody, proceed directly to



6.5 mm Slides					
1X TrueBlack Solution	Stock	Final	1Χ (μl)	2X+ 10% (μl)	4X+ 10% (μl)
TrueBlack Lipofuscin Quencher	-	-	3.5	7.7	15.4
70% Ethanol	-	-	66.5	146.3	292.6
Total	-	-	70.0	154.0	308.0



11 mm Slides					
1X TrueBlack Solution	Stock	Final	1Χ (μl)	2X+ 10% (μl)	4X + 10% (μl)
TrueBlack Lipofuscin Quencher	-	-	7.0	15.4	30.8
70% Ethanol	-	-	133.0	292.6	585.2
Total	-	-	140.0	308.0	616.0

- **c.** Add  $\blacksquare 100 \, \mu l$  or  $\blacktriangle 200 \, \mu l$  1X Blocking Buffer along the side of the wells.
- **d.** Re-apply Visium Slide Seal to the Visium CytAssist Tissue Slide Cassette.
- e. Incubate for 30 min at room temperature.
- f. Peel back slide seal from the cassette and remove all Blocking Buffer from the wells.
- g. Add ■70 μl or ▲140 μl Antibody Solution along the side of the wells. Tap gently to ensure uniform coverage.
- **h.** Re-apply slide seal to the cassette.
- i. Incubate for **60 min** at **room temperature** in the dark.
- j. Peel back slide seal and remove Antibody Solution.
- **k.** Add  $\blacksquare 150 \mu l$  or  $\blacktriangle 300 \mu l$  Wash Buffer along the side of the wells.
- 1. Incubate for **5 min** at **room temperature**.
- m. Remove all Wash Buffer from the wells.
- **n. Repeat** k-m three more times for a total of four washes.

- o. Optional TrueBlack Quenching
  - Add ■70 µl or ▲140 µl 1X TrueBlack Solution along the side of the wells.
  - Incubate for **30 sec** at **room temperature**.
  - Remove all 1X TrueBlack Solution from the wells.
  - Add  $\blacksquare$ 150  $\mu$ l or  $\blacktriangle$ 300  $\mu$ l 1X PBS along the side of the wells.
  - · Remove all PBS from the wells.
  - Repeat PBS washes three more times for a total of four washes.
- p. Remove slide from cassette.



See Tips & Best Practices for removal instructions.

**q.** Gently immerse slide 5x in 1X PBS in 50-ml centrifuge tube. If processing multiple slides, use a separate 50-ml centrifuge tube filled with PBS for each slide.



r. Remove slide from the PBS and proceed immediately to Coverslipping.

DO NOT let the slide dry.

## 1.4B Immunofluorescence Staining - Primary and Secondary Antibodies

- denotes volumes for 6.5 mm gaskets and ▲ denotes volumes for 11 mm gaskets
  - a. Prepare Primary Antibody Solution on ice according to the appropriate table, adding reagents in the order listed. Pipette mix 10x and centrifuge briefly. Maintain on ice.



6.5 mm Gaskets					
Primary Antibody Solution	Stock	Final	1Χ (μl)	2X+ 10% (μl)	4X+ 10% (μl)
Block and Stain Buffer	4X	1X	17.5	38.5	77.0
RNase Inhibitor	40U/μl	2U/µl	3.5	7.7	15.4
Antibody #1	Variable	Variable	Variable	Variable	Variable
Antibody #2 (Optional)	Variable	Variable	Variable	Variable	Variable
Antibody #3 (Optional)	Variable	Variable	Variable	Variable	Variable
Nuclease-free Water	Variable	Variable	Variable	Variable	Variable
Total	-	-	70.0	154.0	308.0



11 mm Gaskets					
Primary Antibody Solution	Stock	Final	1Χ (μl)	2X+ 10% (μl)	4X + 10% (μl)
Block and Stain Buffer	4X	1X	35.0	77.0	154.0
RNase Inhibitor	40U/μl	2U/μl	7.0	15.4	30.8
Antibody #1	Variable	Variable	Variable	Variable	Variable
Antibody #2 (Optional)	Variable	Variable	Variable	Variable	Variable
Antibody #3 (Optional)	Variable	Variable	Variable	Variable	Variable
Nuclease-free Water	Variable	Variable	Variable	Variable	Variable
Total	-	-	140.0	308.0	616.0



Antibody dilution can change depending on the antibody, generally ranging from 1:50 up to 1:1000. Antibody volumes will depend on concentrations determined during antibody optimization.

c. Prepare Secondary Antibody Solution on ice according to the appropriate table, adding reagents in the order listed. Pipette mix 10x and centrifuge briefly. Maintain on ice.



6.5 mm Gaskets					
Secondary Antibody Solution	Stock	Final	1X (µl)	2X+ 10% (μl)	4X+ 10% (μl)
Block and Stain Buffer	4X	1X	17.5	38.5	77.0
RNase Inhibitor	40U/µl	2U/μl	3.5	7.7	15.4
Antibody #1	Variable	Variable	Variable	Variable	Variable
Antibody #2 (Optional)	Variable	Variable	Variable	Variable	Variable
Antibody #3 (Optional)	Variable	Variable	Variable	Variable	Variable
DAPI Solution	100X	1X	0.7	1.5	3.1
Nuclease-free Water	Variable	Variable	Variable	Variable	Variable
Total	-	-	70.0	154.0	308.0



11 mm Gaskets					
Secondary Antibody Solution	Stock	Final	1X (µI)	2X+ 10% (μl)	4X + 10% (μ)
Block and Stain Buffer	4X	1X	35.0	77.0	154.0
RNase Inhibitor	40U/µl	2U/µl	7.0	15.4	30.8
Antibody #1	Variable	Variable	Variable	Variable	Variable
Antibody #2 (Optional)	Variable	Variable	Variable	Variable	Variable
Antibody #3 (Optional)	Variable	Variable	Variable	Variable	Variable
DAPI Solution	100X	1X	1.4	3.1	6.2
Nuclease-free Water	Variable	Variable	Variable	Variable	Variable
Total	-	-	140.0	308.0	616.0



Antibody dilution can change depending on the antibody, ranging from 1:200 up to 1:1000. Antibody volumes will depend on concentrations determined during antibody optimization.

e. Optional - Prepare 1X TrueBlack Solution according to the appropriate table, adding reagents in the order listed. Pipette mix 10x and centrifuge briefly. Maintain at room temperature.



6.5 mm Gaskets					
1X TrueBlack Solution	Stock	Final	1X (µl)	2X+ 10% (μl)	4X+ 10% (μl)
TrueBlack Lipofuscin Quencher	-	-	3.5	7.7	15.4
70% Ethanol	-	-	66.5	146.3	292.6
Total	-	-	70.0	154.0	308.0



11 mm Gaskets					
1X TrueBlack Solution	Stock	Final	1X (µl)	2X+ 10% (µl)	4X + 10% (µl)
TrueBlack Lipofuscin Quencher	-	-	7.0	15.4	30.8
70% Ethanol	-	-	133.0	292.6	585.2
Total	-	-	140.0	308.0	616.0

- **f.** Add ■100 μl or ▲200 μl 1X Blocking Buffer along the side of the wells.
- g. Re-apply Visium Slide Seal on the Visium CytAssist Tissue Slide Cassette.
- **h.** Incubate for **30 min** at **room temperature**.
- i. Peel back slide seal from cassette and remove all Blocking Buffer from the wells.
- j. Add ■70 μl or ▲140 μl Primary Antibody Solution along the side of the wells. Tap gently to ensure uniform coverage.
- **k.** Re-apply slide seal on the cassette.
- **l.** Incubate for **1h** at **room temperature** in the dark.
- m. Peel back slide seal and remove Primary Antibody Solution.
- **n.** Add  $\blacksquare$ **150 \mul or \triangle 300 \mul** 1X Wash Buffer along the side of the wells.
- **o.** Incubate for **5 min** at **room temperature**.
- **p.** Remove 1X Wash Buffer from the wells.
- **q. Repeat** n-p three more times for a total of four washes.
- **r.** Add ■70 µl or ▲140 µl Secondary Antibody Solution along the side of the wells. Tap gently to ensure uniform coverage.
- **s.** Re-apply slide seal on the cassette.
- t. Incubate for **20 min** at **room temperature** in the dark.

- u. Peel back slide seal and remove Secondary Antibody Solution.
- v. Add  $\blacksquare 150 \,\mu l$  or  $\blacktriangle 300 \,\mu l$  1X Wash Buffer along the side of the wells.
- w. Incubate for 5 min at room temperature.
- x. Remove 1X Wash Buffer from the wells.
- **y.** Repeat v-x three more times for a total of four washes.
- **z.** *Optional* TrueBlack Quenching
  - Add ■70 µl or ▲140 µl 1X TrueBlack Solution along the side of the wells.
  - Incubate for **30 sec** at **room temperature**.
  - Remove all 1X TrueBlack Solution from the wells.
  - Add  $\blacksquare 150 \,\mu l$  or  $\blacktriangle 300 \,\mu l$  1X PBS along the side of the wells.
  - · Remove all PBS from the wells.
  - Repeat PBS washes two more times for a total of three washes.
- aa. Remove slide from cassette.



See Tips & Best Practices for removal instructions.

- **ab.** Gently immerse slide 5x in 1X PBS in 50-ml centrifuge tube.
- ac. Remove slide from the PBS and proceed immediately to Coverslipping.



DO NOT let the slide dry.


## 1.5 Coverslipping

- a. Place slide on a flat, clean, non-absorbent work surface. Some residual droplets may remain.
- b. Using a wide-bore pipette tip, add 100 µl Mounting Medium to uniformly cover all tissue sections on slides. Avoid generating bubbles.
- c. Apply the coverslip at an angle on one end of the slide. Slowly lower the coverslip, without introducing bubbles. Allow the mounting medium to spread and settle.
- **d.** Remove any large excess of mounting media by carefully wicking away from the edge of the coverslip with a laboratory wipe. Do not move the coverslip and disturb the tissue.
- e. Once coverslipping is complete, **immediately** proceed with imaging.



DO NOT let the attached coverslip dry.


DO NOT use Cytoseal or nail polish for securing the coverslip. This protocol is designed for use with freshly sectioned tissues.



# 2. Tissue Imaging

#### 2.0 Overview

This chapter provides guidance on imaging tissue slides containing immunofluorescent stained FFPE sections and coverslip removal.



## 2.1 Imaging System Recommendations

The following table shows imaging systems used by 10x Genomics in the development of this protocol. Any equivalent imaging setup can be used as an alternative.

Supplier	Model	Configuration
Thermo Fisher Scientific	EVOS M7000	Inverted
Leica	Aperio Versa 8	Upright
	Leica DMi8	Inverted
MetaSystems	Metafer	Upright
Nikon	Nikon Eclipse Ti2	Inverted
BioTek	Cytation 7	Inverted or Upright
Keyence	Keyence BZX800	Inverted
Olympus	VS200	Upright
Zeiss	Imager.Z2	Upright

Fluorescence Recommended Configuration
Light source (or equivalent) with a wavelength range of 380–680 nm
Monochrome camera (14 bit, 2,424 x 2,424 pixel resolution)
DAPI filter cube (Excitation 392/23, Emission 447/60)
FITC filter cube (Excitation 466/40, Emission 515/30)
TRITC filter cube (Excitation 542/20, Emission 620/52)
Cy5 filter cube (Excitation 618/50, Emission 698/70)
2.18 µm/pixel minimum capture resolution
Exposure times 100 milli sec-2 sec

## 2.2 Imaging


- a. Image all tissue sections individually at the desired magnification using fluorescence imaging settings.
- **b.** Consult the Visium CytAssist Spatial Applications Imaging Guidelines Technical Note (CG000521) for additional information.
- **c.** After imaging, proceed **immediately** to the Coverslip Removal.

## 2.3 Coverslip Removal

- **a.** Dispense **200 ml** 1X PBS in a beaker for every two slides being processed.
- **b.** Immerse the slide sideways/horizontal in the beaker containing **200 ml** 1X PBS with the coverslipped surface fully sideways to prevent the coverslip from dragging across the tissue.
- **c.** Hold the slide in 1X PBS until the coverslip slowly separates away from the slide.



To avoid damaging the tissue sections or causing tissue detachment, DO NOT move the slide up and down, shake forcibly or manually move the coverslip.



- **d.** Gently immerse slide 5x in the 1X PBS to ensure all mounting medium is removed.
- e. Wipe the back of the slide with a laboratory wipe.
- f. Place the slide in a new Visium CytAssist Tissue Slide Cassette. See Tips & Best Practices for assembly instructions.
- **g.** Add  $\blacksquare 100 \, \mu l$  or  $200 \, \mu l$  1X PBS along the side of the wells.
- **h.** Apply a new pre-cut slide seal to the cassette.
- i. Proceed immediately to Visium CytAssist Spatial Gene and Protein Expression User Guide (CG000494).

## **Troubleshooting**

STEP	Notes
Tissue detachment	<ul> <li>Gently add and remove reagents from wells. Forceful removal or addition of reagents can impact adhesion.</li> </ul>
	<ul> <li>Using a validated slide can improve tissue adhesion.</li> </ul>
	<ul> <li>Consult the Troubleshooting section of the Visium CytAssist Spatial Gene and Protein Expression for - FFPE Tissue Preparation Guide (Document CG000660) for more information on sample preparation errors that may cause detachment.</li> </ul>
2.2 Weak or no signal	<ul> <li>Verify that samples were not exposed to light after staining with fluorescent antibodies.</li> </ul>
	<ul> <li>Verify antibody compatibility with decrosslinking conditions.</li> </ul>
	<ul> <li>Verify antibody dilutions. Ensure that antibody optimization is performed.</li> </ul>
	<ul> <li>Verify imaging system filter cubes and wavelength. Ensure that fluorophores and filter cubes match.</li> </ul>
	<ul> <li>Protein of interest may have low expression. Use a positive control.</li> <li>Optimize antibody concentrations on tissues of interest before working with Visium Spatial slides.</li> </ul>
	<ul> <li>If the signals from the primary conjugated antibodies are absent or weak, consider inclusion of secondary antibody staining.</li> </ul>
2.3 High background	<ul> <li>Ensure that samples properly follow all indicated number of washes and incubation times.</li> </ul>
	<ul> <li>Ensurer Blocking Buffer was made correctly and that blocking step was performed accurately.</li> </ul>
	<ul> <li>Verify that samples did not dry out during the staining protocol.</li> <li>Ensure that samples always remain covered in liquid.</li> </ul>
	<ul> <li>To prevent non-specific antibody binding, compare chosen anti- body with antibodies that target the same cell type. If possible, compare staining results to cells known to express higher or lower levels of the target protein.</li> </ul>
	Using TrueBlack may reduce background fluorescence.

## **Document Revision Summary**

**Document Number** CG000659

Title Visium CytAssist Spatial Gene and Protein Expression –

Deparaffinization, Decrosslinking, IF Staining & Imaging for FFPE Tissue

Revision Rev B

Revision Date July 2023

• Clarified how many antibody optimizations each kit supports in a note on

page 3

Added VeritiPro to list of tested thermal cyclers on page 5.

Updated TrueBlack description and part number on page 7.

• Updated TrueBlack dilution instructions on page 26.

LEGAL NOTICE © 2023 10x Genomics, Inc. (10x Genomics). All rights reserved. Duplication and/or reproduction of all or any portion of this document without the express written consent of 10x Genomics, is strictly forbidden. Nothing contained herein shall constitute any warranty, express or implied, as to the performance of any products described herein. Any and all warranties applicable to any products are set forth in the applicable terms and conditions of sale accompanying the purchase of such product. 10x Genomics provides no warranty and hereby disclaims any and all warranties as to the use of any third-party products or protocols described herein. The use of products described herein is subject to certain restrictions as set forth in the applicable terms and conditions of sale accompanying the purchase of such product. A non-exhaustive list of 10x Genomics' marks, many of which are registered in the United States and other countries can be viewed at: www.10xgenomics.com/trademarks. 10x Genomics may refer to the products or services offered by other companies by their brand name or company name solely for clarity, and does not claim any rights in those third-party marks or names. 10x Genomics products may be covered by one or more of the patents as indicated at:www.10xgenomics.com/patents. The use of products described herein is subject to 10x Genomics Terms and Conditions of Sale, available at www.10xgenomics.com/legal-notices, or such other terms that have been agreed to in writing between 10x Genomics and user. All products and services described herein are intended FOR RESEARCH USE ONLY and NOT FOR USE IN DIAGNOSTIC PROCEDURES.

The use of 10x Genomics products in practicing the methods set forth herein has not been validated by 10x Genomics, and such non-validated use is NOT COVERED BY 10X GENOMICS STANDARD WARRANTY, AND 10X GENOMICS HEREBY DISCLAIMS ANY AND ALL WARRANTIES FOR SUCH USE. Nothing in this document should be construed as altering, waiving or amending in any manner 10x Genomics terms and conditions of sale for the Chromium Controller or the Chromium Single Cell Controller, consumables or software, including without limitation such terms and conditions relating to certain use restrictions, limited license, warranty and limitation of liability, and nothing in this document shall be deemed to be Documentation, as that term is set forth in such terms and conditions of sale. Nothing in this document shall be construed as any representation by 10x Genomics that it currently or will at any time in the future offer or in any way support any application set forth herein.

#### **Contact:**

support@10xgenomics.com

10x Genomics 6230 Stoneridge Mall Road Pleasanton, CA 94588 USA

