

Certificate



Certificate no.

CU 72226267 01

License Holder:

10X Genomics, Inc.
6230 Stoneridge Mall Road
Pleasanton CA 94588
USA

Manufacturing Plant:

GES Singapore Pte. Ltd
28 Marsiling Lane
739152
Singapore

Test report no.: USA-JH US22LRD 001

Client Reference: Sural Labha

Tested to:

UL 61010-1:2012 R7.19
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) + A1
UL 61010-2-010:2019
CSA C22.2 No. 61010-2-010:19
UL 61010-2-081:2019
CSA C22.2 No. 61010-2-081:19

Certified Product: Visium CytAssist

License Fee - Units

Model Designation: Visium CytAssist, VPC-LB-0x
(x = 0-9, A-Z, blank; not safety-relevant)

Rated Voltage: AC 100-240V, 50-60Hz
Rated Power: 250W max.
Protection Class: I

Appendix: 1, 1-4

Licensed Test mark:



Date of Issue
(day/mo/yr)

20/05/2022