

Certificate



Certificate no.

CU 72227985 01

License Holder:

10X Genomics, Inc.
6230 Stoneridge Mall Road
Pleasanton CA 94588
USA

Manufacturing Plant:

10X Genomics, Inc.
6230 Stoneridge Mall Road
Pleasanton CA 94588
USA

Test report no.: USA-JHUS22JPUA 001

Client Reference: Sural Labha

Tested to:

UL 61010-1:2012 R7.19
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2 (R2017) + A1
UL 61010-2-010:2019
CSA C22.2 No. 61010-2-010:19
UL 61010-2-081:2019
CSA C22.2 No. 61010-2-081:19

Certified Product: In-Situ Analyzer

License Fee - Units

Model Designation: Xenium Analyzer, 1000529

Rated Voltage:
1) AC 100-240V, 50-60Hz;
2) AC 100-120V, 50-60Hz;
3) AC 200-240V, 50-60Hz

Rated Power:
1) 16-8A;
2) 16A;
3) 8A

Protection Class: I

Appendix: 1, 1-7

Licensed Test mark:



Date of Issue
(day/mo/yr)

03/10/2022